



CERTIFICATE OF ANALYSIS

Product: Polymorphprep™, 1x250 mL
Catalogue No.: 1114683

Lot No.: 00117
Internal No: 10187656
Expiry date: 2020-02

Quality Control Results

Test	Requirement	Result
Density, g/cm ³	1.113 ± 0.001	1.113
Osmolality, mOsm	445 ± 15	444
Endotoxins, EU/mL	< 1.0	< 0.3
Sterility test, Acc. to Ph.Eur	Sterile	Pass

Approved by:

Ulf Næss (UNAS) 2017-04-27 14:56:20 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.