



## CERTIFICATE OF ANALYSIS

**Product:** Lymphoprep

**Product no:** 1114545

**Lot no:** 07119S15

**Expiry date:** 2010-09

**Int. no:** 10132689

**Volume:** 4x250 mL

Test	Result	Requirement
Density	1.076	1.077 +/- 0.001
Osmolality	292	290 +/- 15
Endotoxins	< 0.1	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

**AXIS-SHIELD PoC AS**

2007-11-13

Date

Ruth Elgini

QA Supervisor