



CERTIFICATE OF ANALYSIS

Product: NycoPrep 1.077

Product no: 1114550

Lot no: 07K27S99

Expiry date: 2010-11

Int. no: 10133565

Volume: 4x250 mL

Test	Result	Requirement
Density	1.078	1.077 +/- 0.001
Osmolality	309	295 +/- 15
Endotoxins	< 0.13 EU/mL	< 1.0 EU/mL

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2008-01-09

Date

Ruth Elgin

QA Supervisor