



**CERTIFICATE OF ANALYSIS**  
**for extended Expiry date**

**Product:** Lymphoprep Tube

**Product no:** 1019817

**Lot no:** 10110115

**Expiry date:** 2008-05

**Volume:** 30x2 mL

Test	Result	Requirement
Density	1.078	1.077 +/- 0.001
Osmolality	298	290 +/- 15
Endotoxins	< 1.0 EU/mL	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

**AXIS-SHIELD PoC AS**

2007-06-06  
Date

Kehn Lawson  
QA Supervisor