



CERTIFICATE OF ANALYSIS

Product: Lymphoprep Tube

Product no: 1019818

Lot no: 10129297

Expiry date: 2010-02

Volume: 18x10 mL

Test	Result	Requirement
Density	1.076	1.077 +/- 0.001
Osmolality	300	290 +/- 15
Endotoxins	< 1.0 EU/mL	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2007-06-21
Date

Klein Lawson
QA Supervisor