

CERTIFICATE OF ANALYSIS

Product: Nycodenz[®] AG, 1x500 g
 Catalogue No.: 1002424

 Lot No.: 10169192
 Expiry date: 2016-06

Quality Control Results

Test	Requirement	Result
Water, wt %	≤ 4.0	0.8
Ionic Compounds (as NaCl),%	≤ 0.01	< 0.01
Colour of Solution:		
Absorbance at 400 nm	≤ 0.180	0.145
Absorbance at 420 nm	≤ 0.030	0.007
Absorbance at 450 nm	≤ 0.015	0.002
Assay:		
Iohexol, wt %	98.4 – 101.0	99.5
Microbiological test:		
Bacterial endotoxin (LAL), EU/g	≤ 12	< 0.4

Approved by:

Axis-Shield PoC AS, Quality Assurance,
 Kirsten Jacobsen (KJ) 2014-01-15 13:27:32 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.