



CERTIFICATE OF ANALYSIS

Product: Lymphoprep™ Tube, 18x10 mL
Catalogue No.: 1019818

Lot No.: 10173350
Expiry date: 2017-07

Quality Control Results

Test	Requirement	Result
Density, g/cm ³	1.077 ± 0.001	1.077
Osmolality, mOsm	290 ± 15	280
Endotoxins, EU/mL	< 1.0	< 0.13
Sterility test Acc. to Ph.Eur.	Sterile	Pass

Approved by:

Kirsten Jacobsen (KJ) 2014-09-30 08:22:37 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.