



CERTIFICATE OF ANALYSIS

Product: Lymphoprep™ Tube, 18x10 mL
Catalogue No.: 1019818
Lot No.: 10178253
Expiry date: 2018-04

Quality Control Results

Test	Requirement	Result
Density, g/cm ³	1.077 ± 0.001	1.076
Osmolality, mOsm	290 ± 15	276
Endotoxins, EU/mL	< 1.0	< 0.13
Sterility test Acc. to Ph.Eur.	Sterile	Pass

Approved by:

Anita Pettersen (AP) 2015-09-25 12:12:36 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.