



CERTIFICATE OF ANALYSIS

Product: Lymphoprep™, 4x250 mL
Catalogue No.: 1114545

Lot No.: 12GLS14
Internal No: 10169919
Expiry date: 2016-11

Quality Control Results

Test	Requirement	Result
Density, g/cm ³	1.077 ± 0.001	1.077
Osmolality, mOsm	290 ± 15	279
Endotoxins, EU/mL	< 1.0	< 0.13
Sterility test Acc. to Ph. Eur.	Sterile	Pass

Approved by:

Axis-Shield PoC AS, Quality Assurance,
Kirsten Jacobsen (KJ) 2014-01-15 10:40:11 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.