



## CERTIFICATE OF ANALYSIS

Product: Lymphoprep™, 6x500 mL  
Catalogue No.: 1114547  
  
Lot No.: 12GMS97  
Internal No: 10170917  
Expiry date: 2016-12

### Quality Control Results

Test	Requirement	Result
Density, g/cm <sup>3</sup>	1.077 ± 0.001	1.077
Osmolality, mOsm	290 ± 15	282
Endotoxins, EU/mL	< 1.0	< 0.13
Sterility test Acc. to Ph.Eur.	Sterile	Pass

### Approved by:

Axis-Shield PoC AS, Quality Assurance,  
Kirsten Jacobsen (KJ) 2014-03-14 13:32:06 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.