



## CERTIFICATE OF ANALYSIS

Product: OptiPrep™, 1x250 mL  
Catalogue No.: 1114542  
  
Lot No.: 12HDS16  
Internal No: 10172360  
Expiry date: 2017-04

### Quality Control Results

Test	Requirement	Result
Density, g/cm <sup>3</sup>	1.320 ± 0.001	1.319
Endotoxins, EU/mL	< 1.0	< 0.09
Sterility test. Acc. to Ph.Eur	Sterile	Pass

### Approved by:

Axis-Shield PoC AS, Quality Assurance,  
Kirsten Jacobsen (KJ) 2014-06-11 09:08:13 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.