



CERTIFICATE OF ANALYSIS

Product: Polymorphprep™, 1x250 mL
Catalogue No.: 1114683

Lot No.: 12HLS10
Internal No: 10178545
Expiry date: 2017-11

Quality Control Results

Test	Requirement	Result
Density, g/cm ³	1.113 ± 0.001	1.113
Osmolality, mOsm	445 ± 15	440
Endotoxins, EU/mL	< 1.0	< 0.13
Sterility test, Acc. to Ph.Eur	Sterile	Pass

Approved by:

Anita Pettersen (AP) 2015-08-11 15:44:32 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.