



## CERTIFICATE OF ANALYSIS

Product: Lymphoprep™, 4x250 mL  
Catalogue No.: 1114545  
  
Lot No.: 12IES05  
Internal No: 10182005  
Expiry date: 2018-05

### Quality Control Results

Test	Requirement	Result
Density, g/cm <sup>3</sup>	1.077 ± 0.001	1.077
Osmolality, mOsm	290 ± 15	281
Endotoxins, EU/mL	< 1.0	< 0.13
Sterility test Acc. to Ph. Eur.	Sterile	Pass

### Approved by:

Line Røed (LIROED) 2016-04-07 23:54:28 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.