



CERTIFICATE OF ANALYSIS

Product: Lymphoprep™, 6x500 mL
Catalogue No.: 1114547

Lot No.: 12IFS10
Internal No: 10184704
Expiry date: 2018-06

Quality Control Results

Test	Requirement	Result
Density, g/cm ³	1.077 ± 0.001	1.077
Osmolality, mOsm	290 ± 15	280
Endotoxins, EU/mL	< 1.0	< 0.13
Sterility test Acc. to Ph.Eur.	Sterile	Pass

Approved by:

Farshid Khadivi (FK) 2016-10-10 17:02:12 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.