



CERTIFICATE OF ANALYSIS

Product: Lymphoprep™, 6x500 mL
Catalogue No.: 1114547

Lot No.: 12IIS03
Internal No: 10180618
Expiry date: 2018-09

Quality Control Results

Test	Requirement	Result
Density, g/cm ³	1.077 ± 0.001	1.078
Osmolality, mOsm	290 ± 15	289
Endotoxins, EU/mL	< 1.0	< 0.13
Sterility test Acc. to Ph.Eur.	Sterile	Pass

Approved by:

Anita Pettersen (AP) 2016-01-07 13:39:35 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.