



CERTIFICATE OF ANALYSIS

Product: Lymphoprep™, 6x500 mL
Catalogue No.: 1114547

Lot No.: 12KAS03
Internal No: 10182485
Expiry date: 2019-01

Quality Control Results

Test	Requirement	Result
Density, g/cm ³	1.077 ± 0.001	1.077
Osmolality, mOsm	290 ± 15	283
Endotoxins, EU/mL	< 1.0	< 0.25
Sterility test Acc. to Ph.Eur.	Sterile	Pass

Approved by:

Anita Pettersen (AP) 2016-05-09 09:59:32 (YYYY-MM-DD hh:mm:ss)

This document is electronically signed and is valid without a handwritten signature.