



## CERTIFICATE OF ANALYSIS

**Product:** NycoPrep 1.077

**Product no:** 1114550

**Lot no:** 04K29S16

**Expiry date:** 2007-11

**Int. no:** 10122567

**Volume:** 4x250 mL

| Test       | Result       | Requirement     |
|------------|--------------|-----------------|
| Density    | 1.077        | 1.077 +/- 0.001 |
| Osmolality | 309          | 295 +/- 15      |
| Endotoxins | < 0,13 EU/mL | < 1.0 EU/mL     |

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2005-12-15  
Date

Ruth Elgerin  
QA Supervisor