



## CERTIFICATE OF ANALYSIS

**Product:** NycoPrep 1.077

**Product no:** 1114550

**Lot no:** 06I14S10

**Expiry date:** 2009-09

**Int. no:** 10127011

**Volume:** 4x250 mL

Test	Result	Requirement
Density	1.078	1.077 +/- 0.001
Osmolality	310	295 +/- 15
Endotoxins	< 1.0 EU/mL	< 1.0 EU/mL

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

**AXIS-SHIELD PoC AS**

2006. 11. 16

Date

Kidun Janssen  
QA Supervisor