

CERTIFICATE OF ANALYSIS

 Product: OptiPrep

 Product no: 1114542

 Lot no: 05F15S08

 Expiry date: 2008-06

 Int. no: 10121544

 Volume: 1x250 mL

Test	Result	Requirement
Density	1.320	1.320 +/- 0.001
Osmolalitet	177	170 +/- 15
Endotoxins	< 0.13	< 1.0 EU/mL
Sterility test	Conforms	Passes Ph.Eur

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2005-09-15
Date

Ellen O'Connell
QA Supervisor