



CERTIFICATE OF ANALYSIS

Product: Polymorphprep

Product no: 1114683

Lot no: 05F14S06

Expiry date: 2008-06

Int. no: 10123321

Volume: 1x250 mL

Test	Result	Requirement
Density	1.114	1.113 +/- 0.001
Osmolality	441	445 +/- 15
Endotoxins	<0.13	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation.

AXIS-SHIELD PoC AS

2006.02.14
Date

[Signature]
QA Supervisor