



CERTIFICATE OF ANALYSIS

Product: Lymphoprep

Product no: 1114545

Lot no: 04K29S14A

Expiry date: 2007-11

Int no: 10119057

Volume: 4x250 mL

Test	Result	Requirement
Density	1.076	1.077 +/- 0.001
Osmolality	293	290 +/- 15
Endotoxins	< 0.13 EU/mL	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2005-02-28

Date

Ruth Clayton

QA Supervisor

CERTIFICATE OF ANALYSIS

Product:	<u>Lymphoprep</u>	Product no:	<u>1114544</u>
Lot no:	<u>04K29S14A</u>	Expiry date:	<u>2007-11</u>
Int. no:	<u>10119058</u>		
Volume:	<u>1x250 mL</u>		

Test	Result	Requirement
Density	1.076	1.077 +/- 0.001
Osmolality	293	290 +/- 15
Endotoxins	< 0.13 EU/mL	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2005-02-28
Date

Ruth Elguin
QA Supervisor



CERTIFICATE OF ANALYSIS

Product: Lymphoprep

Product no: 1114547

Lot no: 04L06S02

Expiry date: 2007-12

Int. no: 10119059

Volume: 6x500 mL

Test	Result	Requirement
Density	1.076	1.077 +/- 0.001
Osmolality	291	290 +/- 15
Endotoxins	< 0.13 EU/mL	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2005-02-28

Date

Ruth Elzoni

QA Supervisor