

## CERTIFICATE OF ANALYSIS

**Product:** Lymphoprep **Product no:** 1114547  
**Lot no:** 04L06S02 **Expiry date:** 2007-12  
**Int. no:** 10119156  
**Volume:** 6x500 mL

Test	Result	Requirement
Density	1.076	1.077 +/- 0.001
Osmolality	291	290 +/- 15
Endotoxins	< 0.13 EU/mL	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2005-03-10  
 Date

  
 QA Supervisor