

## CERTIFICATE OF ANALYSIS

**Product:** Lymphoprep **Product no:** 1114545  
**Lot no:** 05F14S05A **Expiry date:** 2008-06  
**Int. no:** 10121158  
**Volume:** 4x250 mL

Test	Result	Requirement
Density	1.076	1.077 +/- 0.001
Osmolality	297	290 +/- 15
Endotoxins	< 0.13	< 1.0 EU/mL
Sterility test	Conforms	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2005-08-22  
Date

Ellen O'Neil  
QA Supervisor