



CERTIFICATE OF ANALYSIS

Product: Lymphoprep

Product no: 1114547

Lot no: 05L15S09

Expiry date: 2008-12

Int. no: 10123348

Volume: 6x500 mL

Test	Result	Requirement
Density	1.076	1.077 +/- 0.001
Osmolality	297	290 +/- 15
Endotoxins	<0.13	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2006.02.14
Date

Adam Lawson
QA Supervisor