



CERTIFICATE OF ANALYSIS

Product: Lymphoprep

Product no: 1114547

Lot no: 06L07S03

Expiry date: 2009-12

Int. no: 10128030

Volume: 6x500 mL

Test	Result	Requirement
Density	1.078	1.077 +/- 0.001
Osmolality	302	290 +/- 15
Endotoxins	< 1.0 EU/mL	< 1.0 EU/mL
Sterility test	Passes	Passes Ph.Eur.

The Product is Produced and Controlled according to internal authorized Production Procedures, Quality Control Procedures and Standard Operation Procedures.

AXIS-SHIELD PoC AS

2007.02.02
Date

Kidri Lawson
QA Supervisor