



## CERTIFICATE OF ANALYSIS

Product: OptiPrep™, 1x250 mL  
Catalogue No.: 1114542  
  
Lot No.: 12HIS20  
Internal No: 10174998  
Expiry date: 2017-09

### Quality Control Results

Test	Requirement	Result
Density, g/cm <sup>3</sup>	1.320 ± 0.001	1.319
Endotoxins, EU/mL	< 1.0	< 1.0
Sterility test. Acc. to Ph.Eur	Sterile	Pass

**Approved by:**